

CBSPD Practice Test Questions and Answers

1. What does CBSPD stand for?

- A) Certification Board for Sterile Processing and Distribution
- B) Central Board for Surgical Procedures and Devices
- C) Clinical Board for Sterilization and Processing
- D) Central Bureau for Sterile Product Development

2. What is the recommended temperature for steam sterilization?

- A) 250°F (121°C)
- B) 270°F (132°C)
- C) Both A and B
- D) 300°F (149°C)

3. Which personal protective equipment is required in the decontamination area?

- A) Surgical mask only
- B) Gloves and gown
- C) Full PPE including fluid-resistant gown, gloves, mask, and eye protection
- D) Hair cover only

4. What is the purpose of a biological indicator in sterilization?

- A) To monitor temperature
- B) To verify sterilization effectiveness
- C) To measure humidity
- D) To check pressure levels

Answers: 1-A 2-C 3-C 4-B

For More CBSPD Questions and Answers FREE, CBSPD Online Prep Training, CBSPD Exam, CBSPD Study Guide, CBSPD Flashcards, CBSPD Quizzes visit:

CBSPD Practice Test