

# CDC Practice Test Questions and Answers

## 1. A dental kiln's thermocouple is used to:

- A) Measure and regulate the vacuum level inside the muffle
- B) Continuously monitor the actual temperature inside the firing chamber
- C) Control the speed of the porcelain tray elevator
- D) Calibrate the shade of the fired porcelain

## 2. What is the primary goal of dental prosthetics design?

- A) To alter the natural bite alignment
- B) To restore oral function, aesthetics, and comfort
- C) To increase tooth sensitivity
- D) To eliminate the need for dental hygiene

## 3. How does digital scanning improve prosthetic fabrication?

- A) It increases the likelihood of errors
- B) It provides high-precision 3D impressions
- C) It eliminates the need for patient input
- D) It reduces customization options

## 4. Metamerism in dental ceramics refers to which phenomenon?

- A) Two restorations appearing to match under one light source but not another
- B) The fluorescence of porcelain under UV light
- C) Color change during firing due to pigment burnout
- D) The shift in translucency as porcelain thickness increases

**Answers: 1-B 2-B 3-B 4-A**

For More CDC Questions and Answers FREE, CDC Online Prep Training,  
CDC Exam, CDC Study Guide, CDC Flashcards, CDC Quizzes visit:

## CDC Practice Test