

CSL Practice Test Questions and Answers

1. Which laboratory safety protocol is most critical when handling potentially infectious blood samples?

- A) Wearing only gloves during sample processing
- B) Following universal precautions and using appropriate PPE
- C) Processing samples in open laboratory areas
- D) Using standard cleaning procedures after handling

2. What is the acceptable temperature range for refrigerated storage of most clinical specimens?

- A) 0°C to 4°C
- B) 2°C to 8°C
- C) 4°C to 10°C
- D) -2°C to 2°C

3. Which quality control measure is essential for ensuring accurate laboratory test results?

- A) Running control samples only when problems occur
- B) Calibrating instruments according to manufacturer specifications
- C) Using expired reagents to save costs
- D) Skipping control checks during busy periods

4. According to CLIA regulations, what is the maximum time limit for reporting critical values to healthcare providers?

- A) Within 30 minutes of discovery
- B) Within 1 hour of discovery
- C) Within 2 hours of discovery
- D) By the end of the business day

Answers: 1-B 2-B 3-B 4-A

For More CSL Questions and Answers FREE, CSL Online Prep Training, CSL Exam, CSL Study Guide, CSL Flashcards, CSL Quizzes visit:

CSL Practice Test